

عنوان آزمایش : Alpha Fetoprotein -CLIA  
عنوان دستگاه : simense

سری ساخت : ۴۰۳۸۲

عنوان کنترل : Biorad II  
شروع کنترل : ۱۳۹۹/۰۵/۲۵

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
84.831	82.154	79.477	76.8	74.123	71.446	68.769	2.677	3.485

تاریخ	مقدار
23/12/1399	73
24/12/1399	75.4
25/12/1399	75
26/12/1399	80
27/12/1399	74.1
28/12/1399	79

تاریخ	مقدار
12/12/1399	75.3
13/12/1399	76.5
14/12/1399	78.2
16/12/1399	78.3
17/12/1399	80
18/12/1399	79.4
19/12/1399	76.8
20/12/1399	76

تاریخ	مقدار
02/12/1399	76
03/12/1399	77.6
04/12/1399	72
05/12/1399	74.5
06/12/1399	75.1
09/12/1399	78
10/12/1399	78.4
11/12/1399	77.8

Mean: 76.655	SD: 2.217	CV%: 2.892
--------------	-----------	------------

