

عنوان آزمایش : CA 15-3-CLIA

عنوان دستگاه : زیمنس

سری ساخت : 1765

عنوان کنترل :

premiunII

شروع کنترل : 1400/09/25


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
51.501	48.634	45.767	42.9	40.033	37.166	34.299	2.867	5.1

تاریخ	مقدار
20/10/1400	42.7
21/10/1400	46.2
22/10/1400	43.6
23/10/1400	45.4
25/10/1400	46.1
26/10/1400	44.4
27/10/1400	45.9
29/10/1400	43

تاریخ	مقدار
07/10/1400	47.3
08/10/1400	47.1
09/10/1400	41.8
11/10/1400	42.9
12/10/1400	42.1
13/10/1400	45.2
14/10/1400	43.8
15/10/1400	43.9
18/10/1400	41.8
19/10/1400	48.7

تاریخ	مقدار
25/09/1400	38.6
27/09/1400	38.3
28/09/1400	36.9
29/09/1400	43.9
30/09/1400	47.4
01/10/1400	44.8
02/10/1400	47.2
04/10/1400	45.6
05/10/1400	43.7
06/10/1400	45.5

Mean: 44.064	SD: 2.826	CV%: 6.414
--------------	-----------	------------



CV% 5.1 SD 2.867 -3SD 34.299 -2SD 37.166 -1SD 40.033 Mean 42.9 1SD 45.767 2SD 48.634 3SD 51.501

چاپ

premiumII

عنوان کنترل

اطلاعات

سری ساخت : 1765/

آرشیو

شروع کنترل : 1400/09/25

مقدار	تاریخ
38.6	1400/09/25
38.3	1400/09/27
36.9	1400/09/28
43.9	1400/09/29
47.4	1400/09/30
44.8	1400/10/01
47.2	1400/10/02
45.6	1400/10/04
43.7	1400/10/05
45.5	1400/10/06
47.3	1400/10/07
47.1	1400/10/08
41.8	1400/10/09
42.9	1400/10/11
42.1	1400/10/12
45.2	1400/10/13
43.8	1400/10/14
43.9	1400/10/15
41.8	1400/10/18
48.7	1400/10/19
42.7	1400/10/20
46.2	1400/10/21
43.6	1400/10/22
45.4	1400/10/23
46.1	1400/10/25

Mean: 44.064 SD: 2.826 CV%: 6.414 60

زیمنس

