

عنوان آزمایش : CA 19-9
عنوان دستگاه : simense

-/سری ساخت : ۱۷۵۲

عنوان کنترل : randox I
شروع کنترل : ۱۳۹۹/۰۵/۳۰

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
27.881	26.885	25.889	24.893	23.897	22.901	21.905	0.996	4.002

تاریخ	مقدار
23/12/1399	26.5
24/12/1399	24.5
25/12/1399	26
26/12/1399	25
27/12/1399	27
28/12/1399	26

تاریخ	مقدار
12/12/1399	25
13/12/1399	25.6
14/12/1399	23.8
16/12/1399	23.87
17/12/1399	25.6
18/12/1399	26
19/12/1399	25
20/12/1399	26.5

تاریخ	مقدار
02/12/1399	25
03/12/1399	24.5
04/12/1399	23.9
05/12/1399	25.3
06/12/1399	23.5
09/12/1399	25.5
10/12/1399	25
11/12/1399	26

Mean: 25.23	SD: 0.953	CV%: 3.777
-------------	-----------	------------

