

عنوان آزمایش : DHEA SO4-CLIA
عنوان دستگاه : abbot

سری ساخت : 33402201

عنوان کنترل : abbot-1
شروع کنترل : 1400/02/01

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
33.833	32.423	31.013	29.603	28.193	26.783	25.373	1.41	4.763

تاریخ	مقدار
22/04/1400	28
23/04/1400	29
24/04/1400	28.6
26/04/1400	27.6
27/04/1400	28.6
28/04/1400	29.8

تاریخ	مقدار
12/04/1400	29
13/04/1400	27
14/04/1400	27.5
15/04/1400	27.9
16/04/1400	28.2
17/04/1400	27.3
19/04/1400	29.8
20/04/1400	27.8
21/04/1400	29.6

تاریخ	مقدار
01/04/1400	27.5
02/04/1400	27
03/04/1400	28
05/04/1400	28.4
06/04/1400	29.4
07/04/1400	27.4
08/04/1400	27.3
09/04/1400	29.7
10/04/1400	27.8

Mean: 28.258	SD: 0.916	CV%: 3.242
--------------	-----------	------------



