

عنوان آزمایش : Digoxin-CLIA

عنوان دستگاه : simense

سری ساخت : ۴۰۳۶۲

عنوان کنترل :

شروع کنترل : ۱۳۹۸/۱۲/۱۱

biorad II

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
2.447	2.373	2.299	2.225	2.151	2.077	2.003	0.074	3.348

تاریخ	مقدار
23/12/1399	2.14
24/12/1399	2.22
25/12/1399	2.3
26/12/1399	2.22
27/12/1399	2.3
28/12/1399	2.3

تاریخ	مقدار
12/12/1399	2.3
13/12/1399	2.22
14/12/1399	2.27
16/12/1399	2.11
17/12/1399	2.27
18/12/1399	2.19
19/12/1399	2.19
20/12/1399	2.13

تاریخ	مقدار
02/12/1399	2.26
03/12/1399	2.21
04/12/1399	2.35
05/12/1399	2.3
06/12/1399	2.27
09/12/1399	2.2
10/12/1399	2.2
11/12/1399	2.3

Mean: 2.239	SD: 0.064	CV%: 2.86
-------------	-----------	-----------

