

عنوان آزمایش : Estriol (Free)-CLIA

عنوان دستگاه : siemens

سری ساخت : 2-40382

عنوان کنترل :

biorad III

شروع کنترل : 1400/06/18

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
10.338	9.965	9.592	9.219	8.846	8.473	8.1	0.373	4.043

تاریخ	مقدار
23/10/1400	9.3
25/10/1400	9.7
26/10/1400	9.5
27/10/1400	9.7
28/10/1400	8.9
29/10/1400	9.5

تاریخ	مقدار
12/10/1400	9.6
13/10/1400	9
14/10/1400	9
15/10/1400	8.8
18/10/1400	9.3
19/10/1400	9
20/10/1400	9.8
21/10/1400	9.5
22/10/1400	10

تاریخ	مقدار
01/10/1400	9.28
02/10/1400	9.06
04/10/1400	9
05/10/1400	9.04
06/10/1400	8.58
07/10/1400	9.18
08/10/1400	8.81
09/10/1400	8.23
11/10/1400	9.1

Mean: 9.203

SD: 0.412

CV%: 4.474

biorad III

عنوان کنترل

اطلاعات

سری ساخت : 2-40382

آرشیو

شروع کنترل : 1400/06/18

مقدار

تاریخ

مقدار	تاریخ
9.28	1400/10/01
9.06	1400/10/02
9	1400/10/04
9.04	1400/10/05
8.58	1400/10/06
9.18	1400/10/07
8.81	1400/10/08
8.23	1400/10/09
9.1	1400/10/11
9.6	1400/10/12
9	1400/10/13
9	1400/10/14
8.8	1400/10/15
9.3	1400/10/18
9	1400/10/19
9.8	1400/10/20
9.5	1400/10/21
10	1400/10/22
9.3	1400/10/23
9.7	1400/10/25
9.5	1400/10/26
9.7	1400/10/27
8.9	1400/10/28
9.5	1400/10/29

Mean: 9.203

SD: 0.412

CV%: 4.474

6σ

CV% 4.043 SD 0.373 -3SD 8.1 -2SD 8.473 -1SD 8.846 Mean 9.219 1SD 9.592 2SD 9.965 3SD 10.338

siemens

