

عنوان آزمایش : Estriol (Free)-CLIA
عنوان دستگاه : siemens

سری ساخت : 40382

عنوان کنترل : biorad II
شروع کنترل : 1399/05/25

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
5.059	4.852	4.645	4.438	4.231	4.024	3.817	0.207	4.675

تاریخ	مقدار
22/09/1400	4.32
23/09/1400	4.77
24/09/1400	4.25
25/09/1400	4.08
27/09/1400	4.48
28/09/1400	4.94
29/09/1400	4.5
30/09/1400	4.18

تاریخ	مقدار
11/09/1400	4.51
13/09/1400	4.59
14/09/1400	4.75
15/09/1400	4.38
16/09/1400	4.48
17/09/1400	4.75
18/09/1400	4.26
20/09/1400	4.25
21/09/1400	4.51

تاریخ	مقدار
01/09/1400	4.57
02/09/1400	4.23
03/09/1400	4.38
04/09/1400	4.69
06/09/1400	4.63
07/09/1400	4.56
08/09/1400	4.82
09/09/1400	5.02
10/09/1400	4.46

Mean: 4.514	SD: 0.239	CV%: 5.293
-------------	-----------	------------



CV% 4.675 SD 0.207 -3SD 3.817 -2SD 4.024 -1SD 4.231 Mean 4.438 1SD 4.645 2SD 4.852 3SD 5.059

biorad II

عنوان کنترل

اطلاعات

سری ساخت : 40382

آرشیو

شروع کنترل : 1399/05/25

مقدار	تاریخ
4.57	1400/09/01
4.23	1400/09/02
4.38	1400/09/03
4.69	1400/09/04
4.63	1400/09/06
4.56	1400/09/07
4.82	1400/09/08
5.02	1400/09/09
4.46	1400/09/10
4.51	1400/09/11
4.59	1400/09/13
4.75	1400/09/14
4.38	1400/09/15
4.48	1400/09/16
4.75	1400/09/17
4.26	1400/09/18
4.25	1400/09/20
4.51	1400/09/21
4.32	1400/09/22
4.77	1400/09/23
4.25	1400/09/24
4.08	1400/09/25
4.48	1400/09/27
4.94	1400/09/28
4.5	1400/09/29

Mean: 4.514 SD: 0.239 CV%: 5.293 6σ

siemens

