

عنوان آزمایش : LH-CLIA

عنوان دستگاه : abbot

سری ساخت : ۳۳۴۰۲۲۰۱

عنوان کنترل :

abbott-1

شروع کنترل : ۱۴۰۰/۰۲/۰۱

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
4.895	4.659	4.423	4.187	3.951	3.715	3.479	0.236	5.641

تاریخ	مقدار
24/03/1400	4.24
25/03/1400	4.1
26/03/1400	4.09
27/03/1400	4.27
29/03/1400	4.18
30/03/1400	4.02
31/03/1400	4.23

تاریخ	مقدار
11/03/1400	3.93
12/03/1400	4.02
13/03/1400	3.8
17/03/1400	3.86
18/03/1400	3.85
19/03/1400	3.99
20/03/1400	4.2
22/03/1400	4.11
23/03/1400	3.96

تاریخ	مقدار
01/03/1400	4.11
02/03/1400	3.93
03/03/1400	4.1
04/03/1400	4.25
05/03/1400	4.14
06/03/1400	4.13
08/03/1400	3.97
09/03/1400	4.27
10/03/1400	3.96

Mean: 4.068	SD: 0.138	CV%: 3.403
-------------	-----------	------------

