

عنوان آزمایش : Phenobarbital
عنوان دستگاه : hplc

سری ساخت : 2-1400

عنوان کنترل : پولد سرم
شروع کنترل : 1400/02/28


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
36.281	34.005	31.729	29.453	27.177	24.901	22.625	2.276	7.729

تاریخ	مقدار
22/10/1400	32.9
25/10/1400	31.6
26/10/1400	31.2
27/10/1400	31.8
28/10/1400	27
29/10/1400	31.8
30/10/1400	30.3

تاریخ	مقدار
12/10/1400	32.5
13/10/1400	29
14/10/1400	27.7
15/10/1400	27.7
18/10/1400	29.4
19/10/1400	26.4
20/10/1400	33.7
21/10/1400	32.9

تاریخ	مقدار
01/10/1400	30.9
02/10/1400	30.1
04/10/1400	30
05/10/1400	30.1
06/10/1400	30.2
07/10/1400	30.1
09/10/1400	30.8
11/10/1400	30.2

Mean: 30.361	SD: 1.904	CV%: 6.273
--------------	-----------	------------





CV% 7.729 SD 2.276 -3SD 22.625 -2SD 24.901 -1SD 27.177 Mean 29.453 1SD 31.729 2SD 34.005 3SD 36.281

جاب

بولد سرم

عنوان کنترل

اطلاعات

سری ساخت : 2-1400

آرشیو

شروع کنترل : 1400/02/28

مقدار تاریخ

30.9	1400/10/01
30.1	1400/10/02
30	1400/10/04
30.1	1400/10/05
30.2	1400/10/06
30.1	1400/10/07
30.8	1400/10/09
30.2	1400/10/11
32.5	1400/10/12
29	1400/10/13
27.7	1400/10/14
27.7	1400/10/15
29.4	1400/10/18
26.4	1400/10/19
33.7	1400/10/20
32.9	1400/10/21
32.9	1400/10/22
31.6	1400/10/25
31.2	1400/10/26
31.8	1400/10/27
27	1400/10/28
31.8	1400/10/29
30.3	1400/10/30

Mean: 30.361

SD: 1.904

CV%: 6.273

6σ

