

عنوان آزمایش : Progesterone CLIA
عنوان دستگاه : ABBOTT

سری ساخت : ۳۳۴۰۲۲۰۱

عنوان کنترل : ABBOTT-1
شروع کنترل : ۱۴۰۰/۰۲/۰۱

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
1.119	1.061	1.003	0.945	0.887	0.829	0.771	0.058	6.147

تاریخ	مقدار
24/03/1400	0.9
25/03/1400	1
26/03/1400	0.95
27/03/1400	1
29/03/1400	1
30/03/1400	0.9
31/03/1400	1

تاریخ	مقدار
11/03/1400	1
12/03/1400	0.95
13/03/1400	1
17/03/1400	1
18/03/1400	0.95
19/03/1400	1
20/03/1400	1.05
22/03/1400	1
23/03/1400	1

تاریخ	مقدار
01/03/1400	1
02/03/1400	0.99
03/03/1400	1
04/03/1400	0.95
05/03/1400	1
06/03/1400	0.9
08/03/1400	1
09/03/1400	1
10/03/1400	0.9

Mean: 0.978	SD: 0.041	CV%: 4.179
-------------	-----------	------------

