

عنوان آزمایش : Rubella Ab (IgM)-CLIA
عنوان دستگاه : لیازون

سری ساخت : 5560400

عنوان کنترل : DIASORIN II
شروع کنترل : 1400/03/09


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
103.385	94.527	85.669	76.811	67.953	59.095	50.237	8.858	11.533

تاریخ	مقدار
22/04/1400	93
23/04/1400	65
24/04/1400	85
26/04/1400	75
27/04/1400	62.5
28/04/1400	82
31/04/1400	63.1

تاریخ	مقدار
12/04/1400	94
13/04/1400	74
14/04/1400	90
15/04/1400	81.7
16/04/1400	60
17/04/1400	85
19/04/1400	84
20/04/1400	93
21/04/1400	94

تاریخ	مقدار
05/04/1400	69
05/04/1400	90
05/04/1400	83.1
05/04/1400	94
06/04/1400	82
07/04/1400	71.2
08/04/1400	90
09/04/1400	70
10/04/1400	85

Mean: 80.624	SD: 11	CV%: 13.644
--------------	--------	-------------



% 11.533 SD 8.858 -3SD 50.237 -2SD 59.095 -1SD 67.953 Mean 76.811 1SD 85.669 2SD 94.527 3SD 103.385

چاپ 🖨

لیازون

